

CONTROL SAMPLES

for Microdialysis Analysers

Ref. 8010203F Rev. Mar 2015

ENGLISH

LOT	T25918	Component	Acceptable control range	
Control Low Ref:8010403		Glucose	1.1 - 1.5 mmol/L	(20 - 27 mg/dL)
		Lactate	0.6 - 1.0 mmol/L	(5 - 9 mg/dL)
		Pyruvate	41 - 69 µmol/L	(0.4 - 0.6 mg/dL)
		Urea	0.9 - 1.6 mmol/L	(6 - 9 mg/dL)
		Glycerol	49 - 81 µmol/L	
Control Elevated Ref: 8010205		Glucose	13.3 - 17.9 mmol/L	(239 - 323 mg/dL)
		Lactate	7.7 - 11.5 mmol/L	(69 - 104 mg/dL)
		Pyruvate	561 - 759 µmol/L	(4.9 - 6.7 mg/dL)
		Urea	12.8 - 17.3 mmol/L	(77 - 104 mg/dL)
		Glycerol	663 - 897 µmol/L	
		Glutamate	102 - 138 µmol/L	

SVENSKA

LOT	T25918	Komponent	Acceptabelt kontrollområde	
Control Low Ref:8010403		Glukos	1,1 - 1,5 mmol/L	(20 - 27 mg/dL)
		Laktat	0,6 - 1,0 mmol/L	(5 - 9 mg/dL)
		Pyruvat	41 - 69 µmol/L	(0,4 - 0,6 mg/dL)
		Urea	0,9 - 1,6 mmol/L	(6 - 9 mg/dL)
		Glycerol	49 - 81 µmol/L	
Control Elevated Ref: 8010205		Glukos	13,3 - 17,9 mmol/L	(239 - 323 mg/dL)
		Laktat	7,7 - 11,5 mmol/L	(69 - 104 mg/dL)
		Pyruvat	561 - 759 µmol/L	(4,9 - 6,7 mg/dL)
		Urea	12,8 - 17,3 mmol/L	(77 - 104 mg/dL)
		Glycerol	663 - 897 µmol/L	
		Glutamat	102 - 138 µmol/L	

DEUTSCH

LOT	T25918	Komponente	Akzeptanzbereich für Kontrollen	
Control Low Ref:8010403		Glukose	1,1 - 1,5 mmol/L	(20 - 27 mg/dL)
		Laktat	0,6 - 1,0 mmol/L	(5 - 9 mg/dL)
		Pyruvat	41 - 69 µmol/L	(0,4 - 0,6 mg/dL)
		Harnstoff	0,9 - 1,6 mmol/L	(6 - 9 mg/dL)
		Glycerin	49 - 81 µmol/L	
Control Elevated Ref: 8010205		Glukose	13,3 - 17,9 mmol/L	(239 - 323 mg/dL)
		Laktat	7,7 - 11,5 mmol/L	(69 - 104 mg/dL)
		Pyruvat	561 - 759 µmol/L	(4,9 - 6,7 mg/dL)
		Harnstoff	12,8 - 17,3 mmol/L	(77 - 104 mg/dL)
		Glycerin	663 - 897 µmol/L	
		Glutamat	102 - 138 µmol/L	

QR Codes

Nominal Control Levels

Service Code

Levels



CLIA Intervals

Service Code

Intervals



<http://www.mdialysis.com/products/clinical-microdialysis-international/products/analyzers/controlsamplesdownload>

