

## Calibration Verification Kit (Ref: 8010920)

for M Dialysis Analyzers

LOT	T26519	Component	Target value	Acceptable range
Level I Ref: 8010924		Glucose	2.59 mg/dL	± 1.5 mg/dl
		Lactate	0.240 mmol/L	± 0.15 mmol/L
		Pyruvate	3.33 µmol/L	± 2 µmol/L
		Glycerol	29.2 µmol/L	± 9 µmol/L
		Glutamate	2.92 µmol/L	± 0.9 µmol/L
Level II Ref: 8010925		Glucose	51.9 mg/dL	
		Lactate	1.44 mmol/L	
		Pyruvate	33.3 µmol/L	± 20 %
		Glycerol	175 µmol/L	
		Glutamate	17.5 µmol/L	
Level III Ref: 8010926		Glucose	104 mg/dL	
		Lactate	2.88 mmol/L	
		Pyruvate	66.7 µmol/L	± 20 %
		Glycerol	350 µmol/L	
		Glutamate	35.0 µmol/L	
Level IV Ref: 8010927		Glucose	207 mg/dL	
		Lactate	5.75 mmol/L	
		Pyruvate	133 µmol/L	± 20 %
		Glycerol	700 µmol/L	
		Glutamate	70.0 µmol/L	
Level V Ref: 8010928		Glucose	415 mg/dL	
		Lactate	11.5 mmol/L	
		Pyruvate	267 µmol/L	± 20 %
		Glycerol	1400 µmol/L	
		Glutamate	140 µmol/L	