

REAGENT Set C

For ISCUS/ISCUS^{flex} Clinical Microdialysis Analyzer

Ref. No. 8002165

Content

1. Reagent: One bottle lyophilized reagent each for Glucose, Lactate and Pyruvate placed in a cassette.
2. Buffer: One bottle 6 mL each for Glucose, Lactate and Pyruvate.
3. Calibrator: One bottle of Calibrator A 6 mL placed in the Reagent Cassette.

Preparation

1. Unscrew the cap with the membrane from the reagent and calibrator bottles, placed in the cassette. Remove and discard the rubber stoppers.
2. Unscrew the cap from the buffer bottles and gently transfer the contents to the corresponding reagent bottle.
3. Fasten the cap with the membrane on the reagent and calibrator bottles in the cassette, without returning the rubber stoppers.
4. Let the reagents and calibrator stand and equilibrate in room temperature for at least 30 minutes prior to use. The contents will be completely mixed when the Reagent Cassette is placed and identified in the analyzer.

Intended use and stability of solution



The Reagent Set is made for use in ISCUS or ISCUS^{flex} Clinical Microdialysis Analyzer. Each Reagent Cassette has a unique code, written on the label, which should be registered when placing it in the analyzer.

The Reagent Set is stable up to expiry date when stored at +2 to +8 °C. Reconstituted reagents are stable for five days in the analyzer. The contents are sufficient for around 310 analyses.

GLUCOSE

Colorimetric method for the quantitative determination of Glucose in Microdialysates.

Measuring principle
Glucose is enzymatically oxidized by glucose oxidase (GOD). The hydrogen peroxide formed reacts with phenol and 4-amino-antipyrine. This reaction is catalyzed by peroxidase (POD) and yields a red-violet colored quinoneimine. The rate of formation is measured photometrically at 530 nm and is proportional to the glucose concentration.



Linear range: 0.1 - 25 mmol/L

	Component in test solution	Concentration
Glucose reagent	4-aminoantipyrine	0.77 mmol/L
	Ascorbate oxidase	>3 kU/L
	Glucose oxidase	>1.5 kU/L
	Peroxidase	>1.5 kU/L
Glucose buffer	Phosphate buffer, pH 7.0	0.1 mol/L
	Phenol	11 mmol/L
	Sodium azide	0.4 g/L



References:

1. Barhem and P. Trinder, Analyst 97 (1972) 142

LACTATE

Colorimetric method for the quantitative determination of Lactate in Microdialysates.

Measuring principle
Lactate is enzymatically oxidized by lactate oxidase. The hydrogen peroxide formed reacts with 4-chlorophenol and 4-amino-antipyrine. This reaction is catalyzed by peroxidase (POD) and yields a red-violet colored quinoneimine. The rate of formation is measured photometrically at 530 nm and is proportional to the lactate concentration.



Linear range: 0.1 - 12 mmol/L

	Component in test solution	Concentration
Lactate reagent	4-aminoantipyrine	0.4 mmol/L
	Lactate oxidase	>500 U/L
	Peroxidase	>500 U/L
	Ascorbate oxidase	>12.0 kU/L
Lactate buffer	PIPES buffer, pH 6.8	100 mmol/L
	4-Chlorophenol	5.4 mmol/L
	Sodium oxalate	7.5 mmol/L
	EDTA-disodium salt	5 mmol/L
	Sodium azide	0.3 g/L



References:

1. N. Shimojo et al. Clin Chem 35 (1989) 1992
2. H.F. Kühnle et al, J.Clin Chem BioChem 15 (1977) 171
3. T.O. Kleine et al, Dtsch Med Wschr 104 (1979) 553

PYRUVATE

Colorimetric method for the quantitative determination of Pyruvate in Microdialysates.

Measuring principle
Pyruvate is enzymatically oxidized by pyruvate oxidase (PyrOx). The hydrogen peroxide formed reacts with N-ethyl-N-(2-hydroxy-3-sulfo-propyl)-m-toluidine and 4-amino-antipyrine. This reaction is catalyzed by peroxidase (POD) and yields a red-violet colored quinoneimine. The rate of formation is measured photometrically at 530 nm and is proportional to the pyruvate concentration.



Linear range: 10 - 300 (10 - 1500) µmol/L

	Component in test solution	Concentration
Pyruvate reagent	4-amino-antipyrine	0.3 mmol/L
	Tiamine pyrophosphate	0.2 mmol/L
	FAD	10 µmol/L
	Pyruvate Oxidase	>0.2 kU/L
	Peroxidase	>0.8 kU/L
Pyruvate buffer	Ascorbate Oxidase	>10 kU/L
	Citrate buffer, pH 6.1	100 mmol/L
	Potassium dihydrogenphosphate	10 mmol/L
	MgCl ₂	10 mmol/L
	TOOS	1.5 mmol/L
	Sodium azide	0.3 g/L

References:

1. B. Sedewitz, et al., J. Bacteriol, 160 (1984) 273-278
2. M. Nawata, et al., Anal Biochem, 190 (1990) 84-87
3. H. Araki and M. Yamada, in: H. U. Bergmeyer (Editor), Methods of Enzymatic Analysis, 3rd ed., Vol 6, Verlag Chemie, Weinheim, 1984

CALIBRATOR A

Calibration values	
Glucose	5.55 mmol/L
Lactate	2.5 mmol/L
Pyruvate	250 µmol/L

Sample material

Microdialysates

WARNING

Do not pipette by mouth. Exercise the normal precautions required for handling laboratory reagents.

The buffer contains Sodium Azide. Avoid ingestion or contact with skin or mucous membranes. In case of skin contact, flush affected area with copious amounts of water. In case of contact with eyes or if ingested, seek immediate medical attention.

Sodium Azide may react with lead and copper plumbing, to form potentially explosive azides. When disposing of such reagents, flush with large volumes of water to prevent azide build up. Exposed metal surfaces should be cleaned with 10 % sodium hydroxide.

For in vitro use only

Symbol declaration



Last day of use



Lot number



Storage temperature



The product meets EU directive for IVD (98/79/EC)

Manufactured by:

M Dialysis AB
Hammarby Fabriksväg 43
SE-120 30 • Stockholm • Sweden
Tel: +46-8-470 10 20
Fax: +46-8-470 10 55
E-mail: info@mdialysis.com
www.mdialysis.com

USA office:

73 Princeton Street
N.Chelmsford • MA 01863 • USA
Phone: +1 978 251 1940, +1-866-868-9236
Fax: +1 978 251-1960
E-mail: usa@mdialysis.com

REAGENT Set C

Till ISCUS/ISCUS^{flex} Clinical Microdialysis Analyser

Ref. Nr. 8002165

Innehåll

1. Reagens: En flaska frystorkat reagens vardera för glukos, laktat och pyruvat placerat i en kasset.
2. Buffert: En flaska innehållande 6 mL vardera för glukos, laktat och pyruvat.
3. Kalibrator: En flaska innehållande 6 mL kalibrator A placerad i reagentkassetten.

Beredning

1. Avlägsna de membranförsedda locken från reagensflaskorna och kalibratorflaskan. Tag bort och kasta gummipropparna.
2. Avlägsna locket från buffertflaskorna och överför försiktigt deras innehåll till motsvarande reagens-flaska.
3. Skruva tillbaka de membranförsedda locken på reagens- och kalibratorflaskorna i kassetten, utan att sätta tillbaka gummiproppen.
4. Låt reagenserna och kalibratorn komma i jämvikt i rumtemperatur under minst 30 minuter innan användning. Innehållet kommer att blandas när reagenskassetten placeras och identifierats i analysinstrumentet.

Avsedd användning och lösningarnas stabilitet


Detta reagensset är avsett att användas i ISCUS eller ISCUS^{flex} Clinical Microdialysis Analyser. Varje reagenskasset har ett unikt nummer på etiketten som måste registreras när kassetten placeras i analysinstrumentet.

Innehållet är stabilt till utgångsdatum vid förvaring i +2 till +8 °C. Tillrett reagens är stabilt i fem dagar i analysinstrumentet. Innehållet räcker till ca 310 bestämningar.

GLUKOS

Kolorimetrisk metod för kvantitativ bestämning av glukos i mikrodialysat.

Mätprincip
Glukos oxideras enzymatiskt i närvaro av glukosoxidas (GOD). Den bildade väteperoxiden reagerar med fenol och 4-aminoantipyrin. Denna reaktion katalyseras av peroxidas (POD) och ger ett röd-violett kinonimin-färgämne. Hastigheten med vilken färgämnet bildas mäts fotometriskt vid 530 nm och är proportionell mot glukoskoncentrationen.



Linjärt område: 0,1 - 25 mmol/L

	Komponent i testlösningen	Koncentration
Glukosreagens	4-aminoantipyrin Askorbatoxid Glukosoxid Peroxidas	0,77 mmol/L >3 kU/L >1,5 kU/L >1,5 kU/L
Glukosbuffert	Fosfatbuffert, pH 7,0 Fenol Natriumazid	0,3 mol/L 11 mmol/L 0,4 g/L


Referenser:

1. Barhem and P. Trinder, Analyst 97 (1972) 142

LAKTAT

Kolorimetrisk metod för kvantitativ bestämning av laktat i mikrodialysat.

Mätprincip
Laktat oxideras enzymatiskt i närvaro av laktatoxididas (LOD). Den bildade väteperoxiden reagerar med 4-klorfenol och 4-aminoantipyrin. Denna reaktion katalyseras av peroxidas (POD) och ger ett röd-violett kinonimin-färgämne. Hastigheten med vilken färgämnet bildas mäts fotometriskt vid 530 nm och är proportionell mot laktatkoncentrationen.



Linjärt område: 0,1 - 12 mmol/L

	Komponent i testlösningen	Koncentration
Laktatreagens	4-aminoantipyrin Laktatoxididas Peroxidas	0,4 mmol/L >500 U/L >500 U/L
Laktatbuffert	Askorbatoxididas PIPES buffert, pH 6,8 4-Klorfenol Natriumoxalat EDTA-dinatrium salt Natriumazid	100 mmol/L 5,4 mmol/L 7,5 mmol/L 5 mmol/L 0,3 g/L


Referenser:

1. N. Shimojo et al. Clin Chem 35 (1989) 1992
2. H.F. Kühnle et al., J.Clin Chem BioChem 15 (1977) 171
3. T.O. Kleine et al., Dtsch Med Wschr 104 (1979) 553

PYRUVAT

Kolorimetrisk metod för kvantitativ bestämning av pyruvat i mikrodialysat.

Mätprincip
Pyruvat oxideras enzymatiskt i närvaro av pyruvatoxididas (PyrOx). Den bildade väteperoxiden reagerar med N-etyl-N-(2-hydroxy-3-sulfo-propyl)-m-toluidin och 4-aminoantipyrin. Denna reaktion katalyseras av peroxidas (POD) och ger ett röd-violett kinondiimin-färgämne. Hastigheten med vilken färgämnet bildas mäts fotometriskt vid 530 nm och är proportionell mot pyruvatkoncentrationen.



Linjärt område: 10 - 300 (10 - 1500) µmol/L





	Komponent i testlösningen	Koncentration
Pyruvatreagens	4-amino-antipyrin Tiaminpyrofosfat 0,2 mmol/L FAD Pyruvatoxididas Peroxidas	0,3 mmol/L 10 µmol/L >0,2 kU/L >0,8 kU/L
Pyruvatbuffert	Askorbatoxididas Citratbuffert, pH 6,1 Kalliumdivätefosfat MgCl ₂ TOOS Natriumazid	>10 kU/L 100 mmol/L 10 mmol/L 10 mmol/L 1,5 mmol/L 0,3 g/L

Referenser:

1. B. Sedewitz, et al., J. Bacteriol, 160 (1984) 273-278
2. M. Nawata, et al., Anal Biochem, 190 (1990) 84-87
3. H. Araki and M. Yamada, in: H. U. Bergmeyer (Editor), Methods of Enzymatic Analysis, 3rd ed., Vol 6, Verlag Chemie, Weinheim, 1984

CALIBRATOR A

Kalibreringsvärdet	
Glukos	5,55 mmol/L
Laktat	2,5 mmol/L
Pyruvat	250 µmol/L

Provmaterial Microdialysat	VARNING
Endast för in vitro användning	Pipettera inte med munnen. Använd de försiktighetsåtgärder som krävs för hantering av laboratoriereagenser. Bufferten innehåller natriumazid. Undvik intag och kontakt med hud eller slemhinnor. Vid hudkontakt, skölj med stora mängder vatten. Vid ögonkontakt eller intag, uppsök läkarhjälp omedelbart. Natriumazid kan reagera med bly- och kopparrör och bildar då högeplosiva azider. Vid avyttring, spola alltid med mycket vatten för att förhindra upplagring av azidsalter i avloppssystemet. Exponerade metalltyr tvättas med 10% natriumhydroxid.
Symbolförklaring	 Sista förbrukningsdag
 Lot-nummer	
 Lagertemperatur	
 Produkten uppfyller EU's direktiv för IVD (98/79/EC)/LVFS 2001:7	

Tillverkad av:
M Dialysis AB
Hammarby Fabriksväg 43
SE-120 30 • Stockholm • Sweden
Tel: +46-8-470 10 20
Fax: +46-8-470 10 55
E-mail: info@mdialysis.com
www.mdialysis.com

USA-kontoret:
73 Princeton Street
N.Chelmsford • MA 01863 • USA
Phone: +1 978 251 1940, +1-866-868-9236
Fax: +1 978 251-1960
E-mail: usa@mdialysis.com