

CONTROL SAMPLES

for Microdialysis Analyzers

USA

LOT	T27524	Component	Acceptable control range
Control Low Ref:8010403		Glucose	19.91 - 26.93 mg/dL
		Lactate	1.20 - 2.00 mmol/L
		Pyruvate	27.50 - 45.83 μmol/L
		Glycerol	48.75 - 81.25 μmol/L
		Glutamate	7.50 - 12.50 μmol/L
Control Elevated Ref: 8010205		Glucose	238.89 - 323.21 mg/dL
		Lactate	7.68 - 11.52 mmol/L
		Pyruvate	187.00 - 253.00 μmol/L
		Glycerol	663.00 - 897.00 μmol/L
		Glutamate	102.00 - 138.00 μmol/L

QR Codes

Nominal Control Levels

Service Code Levels



CLIA Intervals

Service Code



Intervals



<http://www.mdialysis.com/products/clinical-microdialysis-international/products/analyzers/controlsamplesdownload>

