

# CONTROL SAMPLES

## for Microdialysis Analysers

Ref. 8010203G Rev. 2021-10-19

### ENGLISH

LOT	T27811	Component	Acceptable control range	
Control Low Ref:8010403		Glucose	1.1 - 1.5 mmol/L	(20 - 27 mg/dL)
		Lactate	1.2 - 2.0 mmol/L	(11 - 18 mg/dL)
		Pyruvate	27 - 46 µmol/L	(0.2 - 0.4 mg/dL)
		Urea	0.9 - 1.6 mmol/L	(6 - 9 mg/dL)
		Glycerol	49 - 81 µmol/L	
		Glutamate	7.5 - 12.5 µmol/L	
Control Elevated Ref: 8010205		Glucose	13.3 - 17.9 mmol/L	(239 - 323 mg/dL)
		Lactate	7.7 - 11.5 mmol/L	(69 - 104 mg/dL)
		Pyruvate	187 - 253 µmol/L	(1.6 - 2.2 mg/dL)
		Urea	12.8 - 17.3 mmol/L	(77 - 104 mg/dL)
		Glycerol	663 - 897 µmol/L	
		Glutamate	102 - 138 µmol/L	

### SVENSKA

LOT	T27811	Komponent	Acceptabelt kontrollområde	
Control Low Ref:8010403		Glukos	1,1 - 1,5 mmol/L	(20 - 27 mg/dL)
		Laktat	1,2 - 2,0 mmol/L	(11 - 18 mg/dL)
		Pyruvat	27 - 46 µmol/L	(0,2 - 0,4 mg/dL)
		Urea	0,9 - 1,6 mmol/L	(6 - 9 mg/dL)
		Glycerol	49 - 81 µmol/L	
		Glutamat	7,5 - 12,5 µmol/L	
Control Elevated Ref: 8010205		Glukos	13,3 - 17,9 mmol/L	(239 - 323 mg/dL)
		Laktat	7,7 - 11,5 mmol/L	(69 - 104 mg/dL)
		Pyruvat	187 - 253 µmol/L	(1,6 - 2,2 mg/dL)
		Urea	12,8 - 17,3 mmol/L	(77 - 104 mg/dL)
		Glycerol	663 - 897 µmol/L	
		Glutamat	102 - 138 µmol/L	

### DEUTSCH

LOT	T27811	Komponente	Akzeptanzbereich für Kontrollen	
Control Low Ref:8010403		Glukose	1,1 - 1,5 mmol/L	(20 - 27 mg/dL)
		Laktat	1,2 - 2,0 mmol/L	(11 - 18 mg/dL)
		Pyruvat	27 - 46 µmol/L	(0,2 - 0,4 mg/dL)
		Harnstoff	0,9 - 1,6 mmol/L	(6 - 9 mg/dL)
		Glycerin	49 - 81 µmol/L	
		Glutamat	7,5 - 12,5 µmol/L	
Control Elevated Ref: 8010205		Glukose	13,3 - 17,9 mmol/L	(239 - 323 mg/dL)
		Laktat	7,7 - 11,5 mmol/L	(69 - 104 mg/dL)
		Pyruvat	187 - 253 µmol/L	(1,6 - 2,2 mg/dL)
		Harnstoff	12,8 - 17,3 mmol/L	(77 - 104 mg/dL)
		Glycerin	663 - 897 µmol/L	
		Glutamat	102 - 138 µmol/L	

### QR Codes

#### Nominal Control Levels

Service Code



Levels



#### CLIA Intervals

Service Code



Intervals



<https://www.mdialysis.com/product/reagents-for-iscusflex-microdialysis-analyzer-2-2/>

