CONTROL SAMPLES

for Microdialysis Analysers

ENGLISH

LOT T28905	Component	Acceptable control range		
Control Low	Glucose	1.1 - 1.5	mmol/L	(20 - 27 mg/dL)
Ref:8010403	Lactate	1.2 - 2.0	mmol/L	(11 - 18 mg/dL)
	Pyruvate	27 - 46	µmol/L	(0.2 - 0.4 mg/dL)
	Urea	0.9 - 1.6	mmol/L	(6 - 9 mg/dL)
	Glycerol	49 - 81	µmol/L	
	Glutamate	7.5 - 12.5	µmol/L	
Control Elevated	Glucose	13.3 - 17.9	mmol/L	(239 - 323 mg/dL)
Ref: 8010205	Lactate	7.7 - 11.5	mmol/L	(69 - 104 mg/dL)
	Pyruvate	187 - 253	µmol/L	(1.6 - 2.2 mg/dL)
	Urea	12.8 - 17.3	mmol/L	(77 - 104 mg/dL)
	Glycerol	663 - 897	µmol/L	
	Glutamate	102 - 138	µmol/L	

SVENSKA

LOT T28905	Komponent	Acceptabelt kontrollområde		
Control Low	Glukos	1,1 - 1,5	mmol/L	(20 - 27 mg/dL)
Ref:8010403	Laktat	1,2 - 2,0	mmol/L	(11 - 18 mg/dL)
	Pyruvat	27 - 46	µmol/L	(0,2 - 0,4 mg/dL)
	Urea	0,9 - 1,6	mmol/L	(6 - 9 mg/dL)
	Glycerol	49 - 81	µmol/L	
	Glutamat	7,5 - 12,5	µmol/L	
Control Elevated	Glukos	13,3 - 17,9	mmol/L	(239 - 323 mg/dL)
Ref: 8010205	Laktat	7,7 - 11,5	mmol/L	(69 - 104 mg/dL)
	Pyruvat	187 - 253	µmol/L	(1,6 - 2,2 mg/dL)
	Urea	12,8 - 17,3	mmol/L	(77 - 104 mg/dL)
	Glycerol	663 - 897	µmol/L	
	Glutamat	102 - 138	µmol/L	

DEUTSCH

LOT T28905	Komponente	Akzeptanzbereich für Kontrollen		
Control Low	Glukose	1,1 - 1,5	mmol/L	(20 - 27 mg/dL)
Ref:8010403	Laktat	1,2 - 2,0	mmol/L	(11 - 18 mg/dL)
	Pyruvat	27 - 46	µmol/L	(0,2 - 0,4 mg/dL)
	Harnstoff	0,9 - 1,6	mmol/L	(6 - 9 mg/dL)
	Glycerin	49 - 81	µmol/L	
	Glutamat	7,5 - 12,5	µmol/L	
Control Elevated	Glukose	13,3 - 17,9	mmol/L	(239 - 323 mg/dL)
Ref: 8010205	Laktat	7,7 - 11,5	mmol/L	(69 - 104 mg/dL)
	Pyruvat	187 - 253	µmol/L	(1,6 - 2,2 mg/dL)
	Harnstoff	12,8 - 17,3	mmol/L	(77 - 104 mg/dL)
	Glycerin	663 - 897	µmol/L	
	Glutamat	102 - 138	µmol/L	

QR Codes

Nominal Control Levels

Service Code   Levels

CLIA Intervals

Service Code   Intervals

http://www.mdialysis.com/products/clinical-microdialysis-international/products/analyzers/controlsamplesdownload